

Islami Bank Foundation

Purchase & Procurement Department

Head Office, Dhaka-1000

Request for Quotation

Ref. : IBF/PPD/IBHM/Rfq/16/2026
Date : 29/09/2026
Quotation invited by : Islami Bank Foundation
Regarding : Supply of Vein Viewer for Islami Bank Hospital Motijheel, Dhaka.
Terms & Conditions :


- i Information Required in the RFQ :
- Name and address (With Telephone No, Fax No. of the company/firm if any)
 - Related experience of the company.
 - Update copy of Trade Licence, TIN, BIN & Tax return certificate.

ii	Specification/ Particulars of the item/Service	:	S/N.	Descriptions	Qty.																																		
				Item Name: Vein Viewer/ Vein Finder for IBH Motijheel. Model- To be mentioned by Bidders. Brand - To be mentioned by Bidders. Origin- To be mentioned by Bidders. (Without India and Israel) Certification: Must have CE & ISO The device should provide high-precision vein visualization, support deep blood-vessel identification, and allow medical personnel to quickly locate veins for puncture. Technical Specifications <table><tr><th>Particular</th><th>Requirement</th></tr><tr><td>Product Type</td><td>Handheld Vein Finder / Vein Visualization Device</td></tr><tr><td>Application</td><td>Peripheral veins/blood vessels identification for venipuncture and vascular access</td></tr><tr><td>Light Source</td><td>Near-Infrared (NIR) light</td></tr><tr><td>Infrared Wavelength:</td><td>400nm~650nm</td></tr><tr><td>Minimum Detectable Vein Diameter</td><td>≤ 1.0 mm</td></tr><tr><td>Image Accuracy</td><td>≤ 0.2 mm</td></tr><tr><td>Imaging Speed</td><td>≥ 30 frames/sec</td></tr><tr><td>Focusing Distance</td><td>200 ± 20 mm</td></tr><tr><td>Image Size</td><td>Full: 81 × 49 mm; 1/2 and 1/3 size selectable</td></tr><tr><td>Contrast Modes</td><td>Normal, Enhanced and Depth Mode</td></tr><tr><td>Imaging Colors</td><td>Green/Blue, Purple/White, Green/White, Red/White, Black/White</td></tr><tr><td>Brightness Adjustment</td><td>High / Medium / Low</td></tr><tr><td>Battery</td><td>Built-in rechargeable Li-ion battery</td></tr><tr><td>Battery Operating Time</td><td>≥ 4 hours</td></tr><tr><td>Input Power</td><td>100–240 VAC, 50–60 Hz, 0.4 A</td></tr><tr><td>Display</td><td>Built-in colour LCD/LED/TFT display, minimum 3.5" or higher</td></tr></table> Warranty: 01 year from of the date of installation.	Particular	Requirement	Product Type	Handheld Vein Finder / Vein Visualization Device	Application	Peripheral veins/blood vessels identification for venipuncture and vascular access	Light Source	Near-Infrared (NIR) light	Infrared Wavelength:	400nm~650nm	Minimum Detectable Vein Diameter	≤ 1.0 mm	Image Accuracy	≤ 0.2 mm	Imaging Speed	≥ 30 frames/sec	Focusing Distance	200 ± 20 mm	Image Size	Full: 81 × 49 mm; 1/2 and 1/3 size selectable	Contrast Modes	Normal, Enhanced and Depth Mode	Imaging Colors	Green/Blue, Purple/White, Green/White, Red/White, Black/White	Brightness Adjustment	High / Medium / Low	Battery	Built-in rechargeable Li-ion battery	Battery Operating Time	≥ 4 hours	Input Power	100–240 VAC, 50–60 Hz, 0.4 A	Display	Built-in colour LCD/LED/TFT display, minimum 3.5" or higher	
Particular	Requirement																																						
Product Type	Handheld Vein Finder / Vein Visualization Device																																						
Application	Peripheral veins/blood vessels identification for venipuncture and vascular access																																						
Light Source	Near-Infrared (NIR) light																																						
Infrared Wavelength:	400nm~650nm																																						
Minimum Detectable Vein Diameter	≤ 1.0 mm																																						
Image Accuracy	≤ 0.2 mm																																						
Imaging Speed	≥ 30 frames/sec																																						
Focusing Distance	200 ± 20 mm																																						
Image Size	Full: 81 × 49 mm; 1/2 and 1/3 size selectable																																						
Contrast Modes	Normal, Enhanced and Depth Mode																																						
Imaging Colors	Green/Blue, Purple/White, Green/White, Red/White, Black/White																																						
Brightness Adjustment	High / Medium / Low																																						
Battery	Built-in rechargeable Li-ion battery																																						
Battery Operating Time	≥ 4 hours																																						
Input Power	100–240 VAC, 50–60 Hz, 0.4 A																																						
Display	Built-in colour LCD/LED/TFT display, minimum 3.5" or higher																																						

- iii. Other Conditions :
- The bidders must quote rates for items in figures clearly and legibly
 - The bidders must quote rates include with charge regarding VAT, Tax, transport and any others charges.
 - Delivery should be done within 7 days after receiving the work order.

- d. The technical offer and the price offer must be submitted in two separate envelopes with seal.
- iv. Payment system : Bill will be paid after the successful installation & commissioning of the above machine.
- v. Validity of RFQ : 150 (One hundred & Fifty) Days
- vi. Other Information : Other relevant information which the quotation provider/supplier may think necessary must be provided and other information which IBF subsequently requests for must be furnished later on.
- Quotation Deadline : **29.09.2026 to 08.10.2026** during 02.00 pm.
- Reservation of Rights : IBF authority reserves the rights to accept or reject any or all the RFQ (Request for Quotation) without assigning any reasons whatsoever.

With best regards,


Md. Shamsul Alam
Manager & In Charge
Purchase & Procurement Dept.


Muhammad Jasim Uddin
DGM, Head of Operations Wing
Islami Bank Foundation

Bidder's Signature & Seal